

Beamline Hazard Analysis - Beamline X22A

LS-BHA-X22A

Revision: 02

Effective: 10/17/11

Page 1 of 1

The following tasks are included in the beamline safety review. Any operations not listed here, or performed by others than those listed, need additional review by ES&H personnel and/or beamline personnel.

Prepared By: C. Nelson

Approved By: Lori Stiegler

Tasks / Equipment	Performed by:		Controls
	Beamline Staff	Users	
Inert compressed gas	Y	N	<ul style="list-style-type: none"> Compressed gas training – TQ-COMPGAS1 or NSLS User training Transport cylinders using a cylinder cart Properly secure cylinders Wear safety glasses with side shields while connecting lines Leak check gas lines
Magnetic Field from X21 when X21 operating			<ul style="list-style-type: none"> Warning signs No workers with ferro-magnetic implants allowed in area when X21 magnet is running
Cutting/razor blades	Y	Y	<ul style="list-style-type: none"> Be aware of hand positioning Use safety knife when possible Consider using cut resistant gloves for frequent cutting tasks